

Summary of Modifications/Changes in this Update

This Summary of Changes is for information only.
It is not a part of the referenced document, and should not be used for project documentation.

U.S. Department of Veterans Affairs ♦ Office of Construction & Facilities Management

DATE OF THIS VERSION (new)

September 1, 2020

TITLE OF DOCUMENT (new title if applicable):

Dental Compressed-Air Equipment, 22 61 19.74

DATE OF VERSION BEING SUPERSEDED (old):

September 1, 2015

DESCRIPTION OF DOCUMENT (previous title, number, other identifying data):

Dental Compressed-Air Equipment, 22 61 19.74

SUMMARY OF CHANGES IN THIS VERSION:

1. General grammatical changes
2. 1.3 – Updated Applicable Publications
3. 1.4 – Added circulating pumps and steam traps to list of equipment requiring a submittal.
4. 1.5 – Added reference to 01 81 13, Sustainable Construction Requirements in section.
5. 1.6 – Replaced all text to comply with Section 220511.
6. 2.1 – Clarified Dental compressed air control panel shall comply with NFPA 99 and UL 60601-1.
7. 2.1.B.6 – Added requirements for electrical componets to be listed and labeled per NFPA 70.
8. 2.1.F – Added a note that dental vacuum and air panels could be combined in a single equipment control panel.